

PROFICIENCY TESTING POLICY

Revision **05**

April 2018

AUTHOR	REVIEWER	APPROVER
Monirul Hoque Pasha Assistant Director	Md. Mahbubur Rahman Deputy Quality Manager	Md. Monwarul Islam Director General

TABLE OF CONTENTS

SECTION	PAGE
1.0 PURPOSE AND SCOPE	3
2.0 NORMATIVE REFERENCE.....	3
3.0 SPECIFIC POLICY	3
4.0 PROFICIENCY TESTING (PT) COVERAGE REQUIREMENT	4
5.0 PROCEDURES OF PT/ILC ORGANIZED BY BAB.....	4
5.1 For Testing Laboratories	4
5.2 For Calibration Laboratories.....	5

1.0 PURPOSE AND SCOPE

All BAB accredited laboratories are required to participate in Proficiency Testing (PT) or Inter-laboratory Comparison (ILC) Programmes recognized by BAB. This policy is also applicable to inspection bodies, where relevant.

In the context of this document, “laboratories” implies all laboratory types – ie testing, calibration and medical laboratories.

This document details BAB’s policy on proficiency testing with regard to laboratories and where relevant to inspection bodies accredited by BAB.

2.0 NORMATIVE REFERENCE

- 2.1 ISO/IEC 17011 - Conformity assessment -- Requirements for accreditation bodies accrediting conformity assessment bodies
- 2.2 ISO/IEC 17025 - General requirements for the competence of calibration and testing Laboratories
- 2.3 ISO 15189 - Medical Laboratories – Requirements for quality and competence.
- 2.4 ISO/IEC 17020 - Conformity assessment -- Requirements for the operation of various types of bodies performing inspection
- 2.5 ILAC-P9:06/2014 ILAC Policy for Participation in Proficiency Testing Activities

3.0 SPECIFIC POLICY

3.1 Laboratories shall also participate (as far as available and practicable) in international Inter-Laboratory Comparison/ Proficiency Programmes conducted by APLAC, EA or equivalent organizations. BAB provides guidance to the laboratories regarding the proficiency testing programmes through its website, by giving reference to international programmes organised by EPTIS, FAPAS etc. This helps the laboratories in participating in other programmes, in case there are not national programmes available.

3.2 BAB is expected to enforce the requirements of APLAC MR 001: ‘Establishing and Maintaining Mutual Recognition Arrangements amongst Accreditation Bodies’ (Cl. 3.3) and ILAC P9: ‘ILAC Policy for Participation in Proficiency Testing Activities’. Accordingly it will be essential for all its accredited laboratories to demonstrate technical competence of its accredited laboratories by satisfactory participation in International/ Regional/ National Proficiency Testing Programmes, including APLAC.

3.3 The PT / ILC program should be conducted in accordance with ISO/ IEC 17043: Conformity assessment -- General requirements for proficiency testing. For this purpose all alternative techniques covered in ISO/ IEC 17043 shall be acceptable.

3.4 At the time of on-site assessment, assessors shall find out whether the laboratories have participated in International/ Regional/ National recognized PT / ILC programme and whether the necessary corrective action were taken, if in terms of Z-Score, En number etc. the laboratory performance has not been satisfactory.

3.5 Satisfactory performance in a Proficiency Testing Programme shall not be considered only criteria for technical competence.

3.6 BAB shall nominate its accredited/applicant laboratories for participation in APLAC and other international Proficiency Testing programme, as and when they are available. Following sequential preference shall be given to the laboratories.

- Accredited laboratories not having participated in any other APLAC PT programme.

- Accredited laboratories not having participated in BAB PT programme or any other programme recognized by BAB.
- Laboratories having unsatisfactory performance in a past PT programme.
- Laboratories with long period of accreditation.
- Applicant laboratories.

3.7 BAB shall co-ordinate all activities related to PT programme with APLAC/ other international bodies and the participating laboratories.

3.8 BAB sometimes, based on demand, necessity and availability of time, may organize PT/ILC programme by selecting a nodal laboratory according to the procedure described in clause 5.0 of this document.

4.0 PROFICIENCY TESTING (PT) COVERAGE REQUIREMENT

4.1 The PT participation shall be such that each major sub disciplines of a laboratory's scope of accreditation are covered in a cycle of three years. This will of course not apply to those special areas where formal Proficiency Testing Programmes are not available.

4.2 All applicant laboratories are required to successfully participate in PT / ILC in at least one parameter/ type of test/ calibration per discipline at least once, prior to grant of accreditation.

4.3 All accredited laboratories except medical laboratories are required to successfully participate in PT / ILC in at least one parameter/ type of test/ calibration per discipline at least once per year to maintain accreditation.

4.4 The accredited medical laboratories are required to participate, at least once in every quarter (03 months), in the PT / External Quality Assurance Scheme/ ILC, as appropriate.

5.0 PROCEDURES OF PT/ILC ORGANIZED BY BAB

5.1 For Testing Laboratories

- BAB identifies various areas of testing where in Proficiency Testing Programme are required to be initiated. Normally, accredited laboratories/ competent shall be selected as nodal laboratories to conduct the PT programmes. In case, the accredited laboratories are not available for a particular test then non-accredited laboratories with sufficient infrastructure and resources may be considered by BAB to act as nodal laboratories.
- BAB will evaluate each proposal, received from the applicant laboratory w.r.t its merit as per requirements of ISO/ IEC 17043 and select a nodal laboratory.
- The applicant laboratories may be invited to present their proposals in BAB to provide clarification and technical details of the PT program.
- The nodal laboratory in consultation with BAB shall design the programme and their testing specification on PT for the participating laboratories.
- Procurement of materials with appropriate precautions for PT programme shall be the responsibility of the nodal laboratory. The nodal laboratory shall ensure that the source of material is authentic and that the producer certifies the composition & properties of the PT material.
- Nodal laboratories under the guidance of BAB shall conduct homogeneity testing of the materials before dispatching the samples to the participating laboratories.
- The nodal laboratories shall carry out proper packaging and dispatching of the samples to participating laboratories. The sample should be accompanied with the instruction sheet, sample receipt form and result proforma.
- The participating laboratories shall conduct the required test, complete the result sheet and dispatch it to BAB.

- The participating laboratories shall conduct the required test, complete the result sheet and dispatch it to BAB.
- After receiving the results from the participating laboratories, BAB will allot a laboratory code number to each participating laboratory, to maintain confidentiality. The result sheet with code number of all the participating laboratories shall be dispatched by BAB to the nodal laboratory for analysis of test data and statistical calculations for performance evaluation. The analysis and calculations shall be done in consultation with BAB.
- Nodal laboratory shall prepare a report of the specific PT programme. The Nodal laboratory shall also prepare laboratory summary sheets for each participating laboratory indicating the performance score, satisfactory / unsatisfactory (outlier) performance and laboratory code number. The same shall be submitted to BAB.
- The summary sheet of each participating laboratory shall be communicated to the participating laboratories on confidential basis by BAB in form of a report.
- The outlier laboratory shall take necessary corrective actions within two months from the date of receipt of communication from BAB.

5.2 For Calibration Laboratories

- BAB identifies a nodal laboratory either NML-BSTI or BAB accredited laboratories or any competent laboratory determined by BAB to conduct Proficiency Testing (PT) / Inter-Laboratory Comparisons (ILC) as per ISO/ IEC 17043.
- The nodal laboratory coordinating the PT / ILC program regularly interacts with BAB at every stage of ILC.
- The nodal laboratory shall acquire the artifacts required for conducting the PT / ILC program. The nodal body shall assign the value of the artifacts based on its own characterization.
- The nodal laboratory shall submit the technical protocol of the PT / ILC program to BAB for approval and BAB will send a list of accredited / applicant laboratories for participation in PT / ILC program.
- The artifact shall be circulated by nodal laboratory to participating laboratories in sequence following a round robin route.
- Each laboratory shall be required to perform measurement with the Calibration and measurement capability (CMC) following the specific calibration procedure and conditions.
- The laboratory shall send the measurement result to the nodal laboratory within a fixed time allotted to each laboratory.
- At any stage during the round robin test, the nodal laboratory may require the artefacts to be brought back for assessing the stability, if needed. At the end of the round robin test the artifact shall be brought back to the nodal body for final measurement to assess its stability and to detect any change if that is taken place.
- Results received from participating laboratory shall be analyzed statistically by the nodal laboratory. The En number in absolute terms shall be the indicating factor to ascertain the performance of the participating laboratory.
- The nodal laboratory shall submit interim reports to BAB at appropriate intervals, where necessary. Besides above, the nodal laboratory shall submit a final report to BAB on completion of the programme.
- The nodal laboratory shall maintain confidentiality and communicate the results directly to the participating laboratories and BAB.
- The laboratory shall take necessary corrective action in terms of En number.
- BAB shall monitor the progress of corrective actions taken by the Calibration laboratories and shall be in regular correspondence with the participating laboratories in this matter.

END